

m|devices®

IV052203

Standard Bore Extension Set

120cm

2 Luer Activated Valve

2 Slide Clamps



INSTRUCTION FOR USE

DESCRIPTION

m|devices IV Standard Bore Extension set is 120cm in length. This set has two Luer Activated Valves, two slide clamps and has a female luer lock to male luer lock with rotating collar. The set should always be primed prior to connecting to a peripheral catheter or access device on an IV administration set.

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.

CAUTION

- Single use only.
- DO NOT re-sterilise.
- DO NOT store at extreme temperatures and humidity, avoid direct sunlight. Handle with care.
- STERILE (EO), DO NOT use if the package or product has been damaged or contaminated.
- EU Notice: any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and /or patient is established.

STEPS

Ensure aseptic technique is followed at all times.

1. Check that all component connections are secure.
2. Remove protective cap from female luer connector.
3. Connect a pre-filled syringe or IV administration set with sterile fluid to female luer connector.
4. To prime line using a syringe, depress plunger to flush fluid through the set whilst inverting the two Luer Activated Valves and until the fluid drips out of the distal end of the extension set. Check to ensure that all air from extension set tubing and luer activated valves has been expelled. If air bubbles are still visible, continue flushing until completely removed.
5. To prime line by connecting to an already existing primed IV administration set, open roller clamp on IV administration set whilst inverting the two Luer Activated Valves to allow fluid to flow through extension set tubing until it drips out of the distal end of the set. Check to ensure that all air from extension set tubing and Luer Activated Valves has been expelled. If air bubbles are still visible, continue flushing until completely removed. Close roller clamp.
6. Remove protective cap from male luer connector.
7. Connect and secure by inserting male luer connector into access device or peripheral catheter and engage rotating collar.

USING THE LUER ACTIVATED VALVES

1. Remove the protective blue end cap.
2. Holding the housing of the Luer Activated Valve whilst maintaining sterility, insert the luer connector directly into the valve, turning clockwise to secure.
3. When disconnecting, hold the housing of the Luer Activated Valve to prevent disconnection from the line.
4. Apply a new protective cap to Luer Activated Valve after each access.

PRECAUTION

- The security of all connections should be checked once established and monitored during use to prevent disconnection.

m|devices

27 Llewellyn Avenue
Villawood NSW 2163
Australia
T +61(0)2 8718 2800
F +61(0)2 8718 2801



VENOUS ACCESS Extension Sets



BQ Plus Medical Co., Ltd.
No. 18, Che Ye Road
Che Dun Town, Songjiang
201611 Shanghai, China



Prolinx GmbH
Brehmstr. 56
40239 Duesseldorf
Germany

STERILE EO



MD



0197