



## AN062001NS - AN062003NS Oxygen Mask with Reservoir Bag

### INSTRUCTION FOR USE

#### DESCRIPTION

m|devices Oxygen Mask with Reservoir Bag are available in two sizes; paediatric and adult. These masks provide higher concentration of oxygen than is able to be provided with a standard face mask or nasal prongs.

Oxygen concentration: 95%-100% at oxygen flow 5-8L/min.

Elongated mask shape for both sizing options, supplied with 2.1m star lumen oxygen tubing and reservoir bag.

**FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.**

#### CAUTION

- Single use only.
- DO NOT re-sterilise.
- DO NOT store at extreme temperatures and humidity, avoid direct sunlight. Handle with care.
- NON STERILE, DO NOT use if the package or product has been damaged or contaminated.
- EU Notice: any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and /or patient is established.

#### STEPS

1. Remove the product from the packaging.
2. Connect oxygen tubing to mask.
3. Connect distal end of tubing to oxygen source.
4. Set oxygen source to prescribed flow rate, ensuring connections are secure. Check for flow through the device and function of the valves. Before placing the mask on the patient, check that oxygen is flowing through the inlet valve between the reservoir bag and the mask.
5. Allow the reservoir bag to fully inflate before placing on the patient.
6. Place the mask over the patient's nose and mouth with the elastic strap below the ears and around the neck.
7. Secure by gently pulling the ends of the elastic strap.
8. Adjust the nose clip on the nose portion of mask, mold it to fit the face. This precludes leakage of oxygen into the corners of the eyes.

#### PRECAUTION

- Ensure all connections are secured and monitored during use.

#### m|devices

27 Llewellyn Avenue  
Villawood NSW 2163  
Australia  
T +61(0)2 8718 2800  
F +61(0)2 8718 2801



 **Multigate Co., Ltd.**  
7 Lanjiang Road 312000  
Shaoxing China

 **RFX+CARE International A/S**  
Bakkegaardsvej 408  
3050 Humlebaek Denmark

