mldevices

AN090000NS - AN090008NS Guedel Airways



DESCRIPTION

m|devices Guedel Airways are sizing colour coded and non-sterile.

Designed to provide additional support to the airway.

Wide sizing range starting from No. 000 through to No.6.

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.

• Single use only.

- DO NOT re-sterilise.
- DO NOT store at extreme temperatures and humidity, avoid direct sunlight. Handle with care.
- NON STERILE, DO NOT use if the package or product has been damaged or contaminated.
- EU Notice: any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and /or patient is established.

STEPS

- 1. Select appropriate size by measuring from the center of the patient's mouth/incisors to the angle of the jaw.
- 2. Open package and remove Guedel.
- 3. Lubricate the distal end of the Guedel Airway if required, prior to insertion.
- 4. Ensure the inner surface curve is facing up towards the roof of the mouth.
- 5. Once the Guedel is one third inside the mouth, rotate the guedel 180 degrees over the tongue.
- 6. Gently advance the Guedel until the flange of the guedel is positioned against the lips.

m|devices

27 Llewellyn Avenue Villawood NSW 2163 Australia T +61(0)2 8718 2800 F +61(0)2 8718 2801



Multigate Co., Ltd. 7 Lanjiang Road 312000 Shaoxing China EC REP RFX+CARE International A/S Bakkegaardsvej 408 3050 Humlebaek Denmark

🔬 🕲 😿 MD 🛆

