m devices

IR010005 Hysteroscopy Irrigation Set



DESCRIPTION

m|devices Hysteroscopy Irrigation Set is designed for hysteroscopy procedures in providing irrigation to the surgical cavity. Single spike configuration with a manual hand pump to enable the clinician to push fluid through to clear the surgical field to assist with visualisation during the procedure.

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.

CAUTION

- · Single use.
- DO NOT re-sterilise.
- . DO NOT store at extreme temperatures and humidity, avoid direct sunlight. Handle with care.
- . STERILE (EO), DO NOT use if the package or product has been damaged or
- . EU Notice: any serious incident that has occurred in relation to the device should be reported to the manufacture and the competent authority of the Member State in which the user and /or patient is established.

STEPS

- 1. Close click clamp.
- 2. Hang fluid bag and remove protective cover from bag spike outlet.
- 3. Remove protective cap from spike of the irrigation set and insert into fluid bag using aseptic technique.
- 4. To prime the irrigation line, invert flexible pump chamber and open the click clamp until the fluid flows out through the silicone tubing adaptor, checking that all air has been expelled.
- 5. Close click clamp.
- 6. Connect and secure by attaching the silicone tubing adaptor to the inline port on the endoscope/instrumentation sheath.
- 7. Control flow rate using the click clamp until desired flow rate is achieved.
- 8. Squeezing the flexible pump chamber will provide a push of fluid to flush the irrigation set if required to assist.

PRECAUTION

• Use of the flexible pump chamber should only be performed by a qualified clinician, as incorrect use could potentially cause high pressures within the surgical cavity.

mldevices

27 Llewellyn Avenue Villawood NSW 2163 Australia T +61(0)2 8718 2800 F +61(0)2 8718 2801





Multigate Co., Ltd. 7 Lanjiang Road 312000 Shaoxing China



3050 Humlebaek Denmark









