

m|devices®

IV050000

Microbore Extension Set

25cm

INSTRUCTION FOR USE

DESCRIPTION

m|devices IV Microbore Extension Set is 25cm in length. This set is lipid resistant and has a female luer lock to male luer lock with rotating collar. The set should always be primed prior to connecting to a peripheral catheter or access device on an IV administration set.

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.

CAUTION

- Single use only.
- DO NOT re-sterilise.
- DO NOT store at extreme temperatures and humidity, avoid direct sunlight. Handle with care.
- STERILE (EO), DO NOT use if the package or product has been damaged or contaminated.
- EU Notice: any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and /or patient is established.

STEPS

1. Check that the package is well sealed and using aseptic technique, open packaging using peel pouch.
2. Remove protective cap from female luer connector.
3. Connect a pre-filled syringe or IV administration set with sterile fluid to female luer connector.
4. To prime line using a syringe, depress plunger to flush fluid through the set and until the fluid drips out of the distal end of the extension set. Check to ensure that all air from extension set tubing has been expelled and continue flushing until any air bubbles are completely removed.
5. To prime line by connecting to an already existing primed IV administration set, open roller clamp on IV administration set to allow fluid to flow through extension set tubing until it drips out of the distal end of the set. Check to ensure that all air from extension set tubing has been expelled and continue flushing until any air bubbles are completely removed. Close roller clamp.
6. Remove protective cap from male luer connector.
7. Connect and secure by inserting male luer connector into access device or peripheral catheter and engage rotating collar.

PRECAUTION

- The security of all luer connections should be checked once the circuit is established and monitored during use to prevent disconnection.

m|devices

27 Llewellyn Avenue
Villawood NSW 2163
Australia
T +61(0)2 8718 2800
F +61(0)2 8718 2801



 **Multigate Co., Ltd.**
7 Lanjiang Road 312000
Shaoxing China

 **RFX+Care International A/S**
Bakkegaardsvej 408
3050 Humlebaek
Denmark

