



IV110011

3-Way Stopcock

INSTRUCTION FOR USE

DESCRIPTION:

m|devices 3-Way Stopcock allows for 45° rotation offering a tactile feel when the handle locks into the desired position and can be performed single handed. This set is lipid resistant.

Intended to be added to IV administration sets, to deliver simultaneous infusion through the multiple ports, deliver bolus medication and to enable multiple access to the infusion line as required. The 3-Way Stopcock must be primed before connecting to the patient end.

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.

CAUTION

DO NOT use if the package or product has been damaged or contaminated.
DO NOT store at extreme temperatures and humidity, avoid direct sunlight.
DO NOT re-use or re-sterilise

STEPS:

1. Remove 3-Way Stopcock from packaging, maintaining aseptic technique at all times.
2. Remove the cap from male luer connector.
3. Connect to the primed IV line by pulling back the rotating collar on the male luer lock and firmly connect stopcock to connecting component, whilst turning to secure and lock the rotating collar in place.
4. To prime, turn the tap to meet required direction/s for fluid to flow by adjusting the roller regulator on the infusion line to allow fluid to flow through the stopcock and out the first port, allowing air to be expelled.
5. Close primed tap and turn tap to ensure all connection ports have been primed all air has been expelled.
6. If using a prefilled syringe to prime the stopcock, repeat the above steps above ensuring air has been expelled from all connection ports before connecting to the patient connection end.

PRECAUTIONS:

The security of all luer connections should be checked once the circuit is established and monitored during use to prevent disconnection.

1800 023 321

m|devices

27 Llewellyn Avenue Villawood
NSW 2163 Australia T +61(0)2

8718 2800

F +61(0)2 8718 2801

mdevices.com.au

m|devices®