



IV120127- IV121130

Insulin Syringe 1mL and 0.5mL – Fixed Needle



INSTRUCTION FOR USE

DESCRIPTION

m|devices insulin syringe in 1mL and 0.5mL options with fixed needles, are available in two needle lengths: 8mm and 13mm and are offered in 27G, 29G and 30G.

Orange cap for easy identification, along with bold scale markings and a clear barrel.

U-100 Insulin

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.



CAUTION

For single use only
DO NOT use if the package or product has been damaged or contaminated.
DO NOT store at extreme temperatures and humidity, avoid direct sunlight.

STEPS:

1. Holding the insulin syringe with the cap still on, pull back the plunger to the line on your syringe for the required dose.
2. Insert the needle into the rubber top of the insulin vial.
3. Always measure from the top of the syringe plunger. Push the air into the vial.
4. Pull back the plunger to the line on your syringe for the insulin dose prescribed.
5. Tap the syringe with your finger to move air bubbles to the top. Then push gently on the plunger to push all of the air bubbles back into the vial.
6. Remove the syringe from the vial and check for correct dose.
7. Select desired location for the injection and follow facility protocol to prepare skin.
8. Use thumb and index finger (or middle finger) to gently lift (not grab) the skin fold and avoid lifting accompanying muscle.
9. Inject into the raised tissue at 90 degrees. Keep the skin fold raised as the insulin is administered.
10. Once full dose has been administered, remove needle and dispose of syringe immediately into sharps bin/container as per facility protocol.

m|devices

27 Llewellyn Avenue
Villawood NSW 2163 Australia

T +61 (0)2 8718 2900

F +61 (0)2 8718 2801

mdevices.com.au

m|devices[®]