

IV200001

LuerSite Valve

Needleless Access Site

i INSTRUCTION FOR USE

DESCRIPTION

m|devices LuerSite valve - Needleless Access Site is designed for use with luer lock and luer slip connectors. A closed silicone diaphragm stops flow until mechanically opened with a luer connector. Bi-direction flow to allow for injection and aspiration. It is lipid resistant and has clear transparent housing to allow visual inspection when flushing.

FOR USE BY A QUALIFIED CLINICIAN. THE BELOW IS ONLY A SUGGESTION AND FACILITY PROTOCOL MUST BE FOLLOWED FOR ALL CLINICAL PROCEDURES WHERE THIS PRODUCT IS USED.

⚠ CAUTION

- Single use only.
- DO NOT re-sterilise.
- DO NOT store at extreme temperatures and humidity, avoid direct sunlight. Handle with care.
- STERILE (EO), DO NOT use if the package or product has been damaged or contaminated.
- EU Notice: any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and /or patient is established.

STEPS

1. Using aseptic technique, open the packaging and remove the LuerSite Valve.
2. Remove the green cap and according to facility protocol, using a syringe filled with sterile fluid prime the valve to dispel all air.
3. Remove the end protective cap and attach to the corresponding luer connection.
4. Holding the LuerSite Valve maintaining sterility, insert the luer connector directly into the valve, turning clockwise to secure and lock in place.
5. Once the medication/fluid has been injected, Facility protocol should be followed for the clamping sequence if disconnecting from the LuerSite Valve connected to a catheter hub. Apply a new protective cap to LuerSite valve after each access.
6. To disconnect: hold the LuerSite Valve firmly and with a twisting motion, turn the luer anti-clockwise to dis-engage.
7. Flush the LuerSite Valve and connecting catheter with a syringe filled with N/Saline after every use or as per recommended facility protocol.
8. Repeat above steps (4-8) for every instance the LuerSite Valve is accessed and replace with a new protective cap after each use.

PRECAUTIONS

- Only luer connectors are recommended when accessing the LuerSite Valve.
- No sharp needles/instruments are to be used with the LuerSite Valve.
- Ensure all connections are secure and monitored during use.
- Ensure aseptic technique is maintained at all times when accessing the LuerSite Valve and corresponding luer connectors/catheter hubs. Once the cap has been removed from the valve, it should not be allowed to touch any non-sterile surfaces.
- The security of all connections should be checked once the circuit is established and monitored during use to prevent disconnection.

Rev.03

m|devices

27 Llewellyn Avenue
Villawood NSW 2163
Australia
T +61(0)2 8718 2800
F +61(0)2 8718 2801



VENOUS
ACCESS

Needle Free Valves



Multigate Co., Ltd.
4/599 Wuxie Road
Yuecheng District, Shaoxing
Zhejiang, China 312000



RFX+CARE International A/S
Gammel Strandvej 22B
2990 Nivå, Denmark



0123